

SNS COLLEGE OF PHYSIOTHERAPY

Affiliated To The Tamil Nadu Dr. MGR Medical University, Chennai

Coimbatore- 641035


COURSE NAME: ELECTROTHERAPY

SUBJECT CODE: 6281

TOPIC: THERAPEUTIC CURRENT

MEDIUM FREQUENCY CURRENT

- Medium Frequency Current: 1000–10,000 Hz
- Lower skin resistance compared to low frequency currents
- More comfortable for patients
- Used mainly for:
 - Pain relief
 - Muscle stimulation
 - Edema reduction
- Examples:
 - Interferential Current (IFC)
 - Russian Current



INTERFERENTIAL CURRENT(IFC) -DEFINITION

- IFC uses two medium-frequency currents that intersect in tissues
- Interference produces a low-frequency therapeutic current
- Common carrier frequency: 4000 Hz
- Advantages:
 - Deeper penetration
 - Less discomfort
 - Effective pain relief



TYPES OF IFC

- Based on electrode arrangement:
 - 2-Pole IFC
 - 4-Pole IFC
 - Classical IFC
 - Vector IFC





Figure 1.

2- POLE IFC


- Uses two electrodes
- Current is pre-modulated
- Medium frequency carrier current is modulated inside the machine
- Frequencies used:
 - Russian Current – 2000 Hz
 - Medium Frequency Current – 4000 Hz
- Uses:
 - Muscle strengthening
 - Pain relief
 - Easier electrode placement



RUSSIAN CURRENT (2000 HZ)


- Type of 2-pole IFC
- Carrier frequency: 2000 Hz
- Burst-modulated medium frequency current

- Indications:
 - Muscle re-education
 - Muscle strengthening
 - Disuse atrophy
- Advantages:
 - Produces strong muscle contraction
 - Used commonly in sports rehabilitation



4-POLE IFC – Classical Method


- Uses four electrodes
- Two separate currents intersect at right angles
- Carrier frequency: 4000 Hz
- Beat frequency produced in tissues
- (e.g., 4000 Hz & 4100 Hz → 100 Hz)
- Advantages:
 - Greater depth of penetration
 - More effective pain relief



4-POLE IFC – Vector Method

- Modification of classical 4-pole IFC
- Interference field rotates automatically
- Covers a larger treatment area

- Benefits:
 - Uniform stimulation
 - Better pain control
 - Less chance of missing the painful area



COMPARISON – 2-Pole vs 4-Pole IFC


2 -

poles

- The 2-pole IFC uses two electrodes
- In 2-pole IFC, modulation of the current is pre-modulated within the machine.
- The depth of penetration in 2-pole IFC is relatively less, making it suitable for smaller and more superficial areas
- The treatment area covered by 2-pole IFC is small.
- 2-pole IFC is comfortable,


4- poles

- Whereas the 4-pole IFC uses four electrodes placed in a criss-cross pattern.
- In 4-pole IFC, modulation occurs within the body tissues due to interference of two medium-frequency currents.
- Whereas 4-pole IFC provides greater depth of penetration, allowing effective treatment of deeper tissues.
- 4-pole IFC can treat a larger area.



SUMMARY & CLINICAL IMPORTANCE

- Medium Frequency Currents are comfortable and effective
- 2-Pole IFC: Simple, superficial, muscle work
- 4-Pole IFC: Deeper penetration, pain relief
- Russian Current: Best for muscle strengthening
- Widely used in orthopaedic & neurological rehab



Thank

you